

KUSAM-MECO®

An ISO 9001:2008 Company

**HIGH VOLTAGE 10KV
DIGITAL INSULATION RESISTANCE TESTER****Model- KM 2804IN - 10KV DC****SAFETY :**

CE marking according to the Low Voltage Directive (2006/95/EC) and EMC Directive (2004/108/EC, 92/31/EEC, & 93/68/EEC) and found to comply with the essential requirement of the directives.

IEC/EN 61010-1: 2001. EN 61326-1, EN 55011+A1+A2, EN 61000-3-2+A2, EN 61000-3-3+A1+A2, EN 61000-4-2+A1+A2, EN 61000-4-3+A1, EN 61000-4-4, EN 61000-4-5+A1, EN 61000-4-6+A1, EN 61000-4-8+A1, EN 61000-4-11, EN61000-4-6+A1, EN61000-4-11.

ELECTRICAL SPECIFICATIONS: KM 2804IN

Test Voltage (DCV)	1 KV	2.5 KV	5 KV	10 KV
Measuring Ranges (Auto-Range)	0-50G	0-125G	0-250G	0-500G
Accuracy	0-50G : ± 3% of reading 50-500G : ± 5% of reading			
Pollution Degree	2			
Altitude	upto 2000 meter			
Operating Temperature & Humidity	0 ~ 40°C; 80% max. R.H.			
Storage Temperature & Humidity	-10°C ~ 65°C; 80% R.H.			
Power Source	1.5V (AA) x 8 Alkaline Batteries			
Weight (battery included)	930g Approx.			
Dimensions	170(L)x120(W)x95(D) mm.			
Accessories	Test leads-AL-50 (3 meter), AL-30 (1.25 meter), AL-30C (1.25 meter), Shoulder belt (BET-2800), Instruction Manual, Batteries			



LCD DISPLAY



AL-50 TestLead standard



AL-30 TestLead



All Specifications are subject to change without prior notice

KUSAM-MECO®

An ISO 9001:2008 Company

G-17, Bharat Industrial Estate, T. J. Road, Sewree (W), Mumbai - 400 015. INDIA.
Sales Direct.: 022 -24156638, **Tel. :** 022-241224540, 24181649, **Fax :** 022 - 24149659
Email : kusam_meco@vsnl.net,
Website : www.kusamelectrical.com